



Shuster Laboratories, Inc.
85 John Road
Canton, MA 02021
Phone: 781-821-2200
Fax: 781-821-9266
www.shusterlabs.com

TAQA®
(Technically Advanced Quality Assurance)
Evaluation Report

Mr. Mahmoud M. Sales
SCI International, Inc.
5902 Enterprise Court
Fredrick, MD 21703

Folder #: 0807894
Date: 4/30/2008
Received Date: 4/9/08
Page: 1 of 3

Product:	New Choice Ovulation Test Strips	UPC Code:	--
Size:	Sample	Reference #:	Partial Prepaid Chk #2637
Supplier:	SCI International	Inventory ID #:	143368
Code:	Lot 3-L03268 EXP 04/2010	Lab Sample #:	0807894001
Purpose:	TAQA EVALUATION vs. ANSWER		--

National Reference Brand: Answer Daily Ovulation Tracker; Lot Code: CV7268 097J41 EXP 08-09 - provided by SCI International

Overall Rating: 10.0 Sample submitted meets TAQA quality criteria for the parameters examined.
Bench top performance is comparable to Answer.

Overall Quality Rating *

- 9.0 - 10.0 Sample tested is **Comparable to the National Brand; meets TAQA quality criteria**
- 7.5 - < 9.0 Sample tested is **Marginally Comparable to the National Brand; does not meet all TAQA quality criteria**
- < 7.5 Sample tested is **Not Comparable to the National Brand; does not meet TAQA quality criteria**

Maryann W. Sullivan
Client Services Manager

Determination		Answer Daily Ovulation Tracker	SCI Sample of New Choice Ovulation Test Strips	Comments
Performance-Bench Top	Positive Result	Obvious positive result 3 min 14 sec after 10 second dip; result remained 10 minutes after dip; meets claim of within 5 minutes	Obvious positive result 1 min 29 sec after 10 second dip; result remained 10 minutes after dip; no claim provided	
	Negative Result	Obvious negative result 4 min 35 sec after 10 second dip; result remained 10 minutes after dip; meets claim of within 5 minutes	Obvious negative result 46 seconds after 10 second dip; result remained 10 minutes after dip; no claim provided	
Physical	Stripe Dimensions, inches	--	--	
	Length	3-1/2	3	
	Width	3/16	1/8 *	
	Thickness	0.039	0.034	
	Strip Description	White plastic "stick" with a green "dipping" line and a test/control panel for result; pink tab on holding end printed with "Ovulation"	White plastic "stick" with a black "dipping" line and a blank panel for result; green tab on holding end	
	Ease of Opening/Closing Container	Acceptable	Acceptable (sample package)	

Mr. Mahmoud M. Sales
SCI International, Inc.
 5902 Enterprise Court
 Fredrick, MD 21703

Folder #: 0807894
 Date: 4/30/2008
 Received Date: 4/9/08
 Page: 3 of 3

		Answer Daily Ovulation Tracker	SCI Sample of New Choice Ovulation Test Strips	Comments
Physical	Product Workmanship	Acceptable	Acceptable	
	Package Workmanship	Acceptable	Sample package	
	Package Description	Opaque/white opaque plastic tube with opaque/white plastic cap; boxed	Opaque/white opaque plastic tube with opaque/white plastic cap (sample package)	